


## Declaration of Conformity

PRODUCT IDENTIFICATION		
Product Name	Model Number	
DiagnosticPRO Advantage	16323	
DosimetryPRO Advantage	17500	
CADPRO Advantage	17032, 19094	
DiagnosticPRO Edge	19523	
Sierra Advantage	14048, 14050	
Dental Film Digitizer	18800	
TeleradPRO Edge HD-CCD	19510, 19511	
MANUFACTURER		
Name of Company	Telephone	Representative
VIDAR Systems Corporation, a 3D Systems Company	365 Herndon Parkway Herndon, VA 20170 USA	703-471-7070 (phone) 703-471-1165 (fax)
AUTHORIZED REPRESENTATIVE		
Name of Company	Address	Telephone/email
Emergo Europe	Molenstraat 15 2513 BH The Hague, Netherlands	+31.70.345.8570 - phone +31.70.346.7299 - fax service@emergogroup.com
CONFORMITY ASSESSMENT		
Device Classification	CE Certificate #	Route to Compliance
Class 1, Rule 12 Non-sterile, non-measuring	Not Applicable for Class I Non-sterile, non-measuring	Annex VII of MDD 93/42/EEC Council Directive
HARMONIZED STANDARDS		
<p>The product identified above complies with the requirements of the EU Directives below by meeting the following standards:</p> <p style="margin-left: 40px;">IEC 60601-1: (3<sup>rd</sup> edition):2005 + CORR. 1(2006) + CORR. 2(2007)</p>		

VIDAR Systems Corporation declares that the above mentioned products meet the provision of the Council Directive 93/42/EEC for Medical Devices and Directive 2011/65/EU on the restriction of the use of certain hazardous substances in electrical and electronic equipment. All supporting documentation is retained at the premises of the manufacturer.

*I am authorized as a company representative to sign as authority on this Declaration of Conformance:*

  
 \_\_\_\_\_  
 Brian Loch, Director of R & D

4/27/2016  
 \_\_\_\_\_  
 Date