

Manufacturer Disclosure Statement for Medical Device Security – MDS²

SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Digitizer	VIDAR Systems Corporation	17110-001 Rev F	11/14/2012
Device Model	Software Revision	Software Release Date	
ClinicalExpress	6.1	11/14/2012	
Manufacturer or Representative Contact Information:	Company Name Representative Name/Position	Manufacturer Contact Information	
	VIDAR Systems Corporation Robert May/Manager Global Partner Support	VIDAR Systems Corporation: Phone:703-471-7070 Email bob.may@3dsystems.com	

<u>MANAGEMENT OF ELECTRONIC PROTECTED HEALTH INFORMATION (ePHI)</u>	<u>Yes</u>	<u>No</u>	<u>N/A</u>	<u>Note #</u>
1. Can this device transmit or maintain electronic Protected Health Information (ePHI)?.....	Yes			_____
2. Types of ePHI data elements that can be maintained by the device:				
a. Demographic (e.g., name, address, location, unique identification number)?.....	Yes			_____
b. Medical record (e.g., medical record #, account #, test or treatment date, device identification number)?.....	Yes			_____
c. Diagnostic/therapeutic (e.g., photo/radiograph, test results, or physiologic data with identifying characteristics)?.....	Yes			_____
d. Open, unstructured text entered by device user/operator?.....	Yes			_____
3. Maintaining ePHI - Can the device				
a. Maintain ePHI temporarily in volatile memory (i.e., until cleared on by power-off or reset)?.....	Yes			_____
b. Store ePHI persistently on local media?.....	Yes			_____
c. Import/export ePHI with other systems?.....	Yes			_____
4. Mechanisms used for the transmitting, importing/exporting of ePHI – Can the device				
a. Display ePHI (e.g., video display)?.....	Yes			_____
b. Generate hardcopy reports or images containing ePHI?.....	No			_____
c. Retrieve ePHI from or record ePHI to removable media (e.g., disk, DVD, CD-ROM, tape, CF/SD card, memory stick)?.....	Yes			_____
d. Transmit/receive or import/export ePHI via dedicated cable connection (e.g., IEEE 1073, serial port, USB, FireWire)?.....	No			_____
e. Transmit/receive ePHI via a network connection (e.g., LAN, WAN, VPN, intranet, Internet)?.....	Yes			_____
f. Transmit/receive ePHI via an integrated wireless connection (e.g. WiFi, Bluetooth, infrared)?.....	No			_____
g. Other?.....				_____

<u>ADMINISTRATIVE SAFEGUARDS</u>	<u>Yes</u>	<u>No</u>	<u>N/A</u>	<u>Note #</u>
5. Does manufacturer offer operator and technical support training or documentation on device security features?.....	Yes			_____
6. What underlying operating system(s) (including version number) are used by the device?.....	See Notes			_____

<u>PHYSICAL SAFEGUARDS</u>	<u>Yes</u>	<u>No</u>	<u>N/A</u>	<u>Note #</u>
7. Are all device components maintaining ePHI (other than removable media) physically secure (i.e. cannot remove without tools)?.....	See Notes			_____
8. Does the device have an integral data backup capability (i.e., backup onto removable media like tape, disk)?.....	No			_____
9. Can the device boot from uncontrolled or removable media (i.e., a source other than an internal drive or memory component)?.....	No			_____

<u>TECHNICAL SAFEGUARDS</u>	<u>Yes</u>	<u>No</u>	<u>N/A</u>	<u>Note #</u>
10. Can software or hardware not authorized by the device manufacturer be installed on the device without the use of tools?.....	Yes			_____
11. Can the device be serviced remotely (i.e., maintenance activities performed by service person via network or remote connection)?.....	Yes			_____
a. Can the device restrict remote access to specific devices or network locations (e.g., specific IP addresses)?.....	Yes			_____
b. Can the device provide an audit trail of remote-service activity?.....	No			_____
c. Can security patches or other software be installed remotely?.....	Yes			_____
12. Level of owner/operator service access to device operating system: Can the device owner/operator				
a. Apply device manufacturer-validated security patches?.....	Yes			_____
b. Install or update antivirus software?.....	Yes			_____
c. Update virus definitions on manufacturer-installed antivirus software?.....	Yes			_____
d. Obtain administrative privileges (e.g. access operating system or application via local root or admin account)?.....	Yes			_____
13. Does the device support user/operator specific username and password?.....	Yes			_____
14. Does the system force reauthorization after a predetermined length of inactivity (e.g., auto logoff, session lock)?.....	Yes			_____

Manufacturer Disclosure Statement for Medical Device Security – MDS²

SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Digitizer	VIDAR Systems Corporation	17110-001 Rev F	11/14/2012
Device Model	Software Revision	Software Release Date	
ClinicalExpress	6.1	11/14/2012	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	VIDAR Systems Corporation	VIDAR Systems Corporation;; Phone 703-471-7070 Email: bob.may@3dsystems.com	
	Representative Name/Position		
	Robert May/Manager Global Partner Support		

15. Events recorded in device audit trail (e.g., user, date/time, action taken): Can the audit trail record.....
- a. Login and logout by users/operators?..... Yes
 - b. Viewing of ePHI?..... Yes
 - c. Creation, modification or deletion of ePHI?..... Yes
 - d. Import/export or transmittal/receipt of ePHI?..... Yes
16. Does the device incorporate an emergency access ("break-glass") feature that is logged?..... No
17. Can the device maintain ePHI during power service interruptions?..... No
18. Controls when exchanging ePHI with other devices:.....
- a. Transmitted only via a point-to-point dedicated cable?..... No
 - b. Encrypted prior to transmission via a network or removable media?..... No
 - c. Restricted to a fixed list of network destinations..... No
19. Does the device ensure the integrity of the ePHI data with implicit or explicit error detection/correction technology?..... Yes

Other Security Considerations

Manufacturer Disclosure Statement for Medical Device Security – MDS²

SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Digitizer	VIDAR Systems Corporation	17110-001 Rev F	11/14/2012
Device Model	Software Revision	Software Release Date	
ClinicalExpress	6.1	11/14/2012	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	VIDAR Systems Corporation	VIDAR Systems Corporation;; Phone 703-471-7070	
	Representative Name/Position	Email: bob.may@3dsystems.com	
	Robert May/Manager Global Partner Support		

SECTION 2

EXPLANATORY NOTES (from questions 1 - 19)

IMPORTANT: Refer to Section 2.2.2 of this standard for the proper interpretation of information requested in this form

- 5. We support Windows Operating Systems: XP SP3 and Windows 7 SP1 both 32 and 64 bit OS.
- 7. Hardware is supplied by Third Party Company. VIDAR only supplies the Software to be installed on the Hardware. This question should be directed to the Hardware provider.
- 19. Windows TCP/IP is the only DICOM supported method we use for transmitting information.